

# Anti-Human CD49b (TEA1/41)



PURE	49BPU	1 mg	1 mg/ml	
FITC	49BF-100T	100 test	20 µL/test	2 mg/ml
PE	49BPE-100T	100 test	20 µL/test	2 mg/ml
APC	49BA-100T	100 test	20 µL/test	2 mg/ml

**RUO**

## 1. PRODUCT DESCRIPTION

**Clone:** TEA1/41;  
**Isotype:** IgG1;  
**Tested application:** flow cytometry;  
**Immunogen:** The anti-CD49b monoclonal antibody derives from TNF activated HUVEC cells;  
**Species reactivity:** Human;  
**Storage instruction:** store in the dark at 2-8 °C;  
**Storage buffer:** aqueous buffered solution containing protein stabilizer and 0.09% sodium azide (NaN<sub>3</sub>);  
**Recommended usage:** Immunostep's CD49b, clone TEA1/41 is a monoclonal antibody intended for the identification and enumeration of Integrin alpha-2 protein using flow cytometry. This reagent is effective for direct immunofluorescence staining of human tissue for flow cytometric analysis using 1 test for 10<sup>6</sup> cells;  
**Presentation:** liquid;  
**Source:** Supernatant proceeding from an in vitro cell culture of a cell hybridoma;  
**Purification:** Affinity chromatography;  
**Other names:** Integrin alpha-2, CD49 antigen-like family member B, Collagen receptor, Platelet membrane glycoprotein Ia, GPIa, VLA-2 subunit alpha, BR, VLAA2;  
**Gene ID:** 3673;  
**Molecular weight:** 170 kDa.

## 2. ANTIGEN DETAILS

**Large description:** The monoclonal antibody is directed against the CD49b-antigen (GPIa or VLA-2 alpha-chain), which can form distinct complexes with either the CD29-antigen (GPIIa or VLA beta-chain), resulting in the VLA-2 (alpha-2 beta- 1) complex, which is expressed on human platelets. The monoclonal antibody reacts with platelets, long-term cultivated T lymphocytes and activated T lymphocytes. In immunohistology the monoclonal antibody reacts with thymocytes, epithelial cells of a variety of tissues, peripheral nerves, fibroblasts, osteoclasts, glomerular mesangium and most non-haemopoietic adherent cell lines.<sup>(1-3)</sup>

## 3. WARRANTY

Warranted only to conform to the quantity and contents stated on the label or in the product labelling at the time of delivery to the customer. Immunostep disclaims hereby other warranties.

Immunostep's sole liability is limited to either the replacement of the products or refund of the purchase price.

## 4. ADDITIONAL INFORMATION

For research use only. Not for diagnostic use.

Not for resale. Immunostep will not be responsible of violations that may occur with the use of this product. Any use of this product other than the specified in this document is strictly prohibited.

Unless otherwise indicated by Immunostep by written authorization, this product is intended for research only and is not to be used for any other purpose, including without limitation, for human or animal diagnostic, therapeutic or commercial purposes.

Please, refer to [www.immunostep.com](http://www.immunostep.com) technical support for more information.

## 5. PROTOCOL

### ■ Direct Immunofluorescence Cell Surface Staining Protocol

1. Transfer 100 µl (10<sup>6</sup> cells/test) of the sample to a 12 x 75 mm polystyrene test tube.
2. Add the suggested volume indicated on the antibody vial to the 12x75 mm cytometer tube.
3. Mix well and incubate in the dark at room temperature at 4 °C for 30 minutes or at room temperature (20-25 °C) for 15 minutes.
4. After the incubation period, add 1,5 ml of an erythrocyte-lysing solution and mix. Incubate at room temperature in the darkness (the blood should be well mixed with the lysing solution).
5. Centrifuge tubes at 540xg for 5 minutes. The supernatant is removed with a Pasteur pipette or with a vacuum pump.
6. Resuspend and wash with 3-5 mL of PBS at 540xg for 5 min.
7. After removing the supernatant and resuspending the cell pellet, add 300 µL of PBS and acquire on the flow cytometer are recorded.
8. Analyse on a flow cytometer or store at 2- 8 °C in the dark until analysis. Samples can be run up to 24 hours after lysis.








### ■ Indirect Immunofluorescence Cell Surface Staining Protocol

1. Transfer 100 µl (10<sup>6</sup> cells/test) of the sample to a 12 x 75 mm polystyrene test tube
2. Add purified reagent according to manufacturer's recommendation and mix gently with a vortex mixer.
3. Incubate in the dark at room temperature at 4 °C for 30 minutes or at room temperature (20-25 °C) for 15 minutes.
4. Add 2 mL 0.01 mol/L PBS (It betters that it containing 2% bovine serum albumin) and resuspend the cells by using a vortex mixer. Centrifuge at 540xg for 5 min in order to remove the McAb not bound to its antigen.
5. Add a secondary conjugated reagent with some fluorochrome and mix. Incubate at room temperature for 15 min in the dark. The absence of light is necessary as the fluorochrome is photoinstability.
6. After the incubation period, add 1,5 ml of an erythrocyte-lysing solution and mix. Incubate at room temperature in the darkness (the blood should be well mixed with the lysing solution). Centrifuge at 540xg for 5 minutes. The supernatant is removed with a Pasteur pipette or with a vacuum pump.
7. Resuspend and a made a final wash with 3-5 mL of PBS at 540xg for 5 min.
8. After removing the supernatant and resuspending the cell pellet, add 300 µL of PBS and acquire on the flow cytometer are recorded.
9. Analyse on a flow cytometer or store at 2- 8 °C in the dark until analysis. Samples can be run up to 24 hours after lysis.

## 6. REFERENCES

1. Mazurov AV, Vinogradov DV, Kabaeva NV, Antonova GN, Romanov YA, Vlasik TN, et al. A monoclonal antibody, VM64, reacts with a 130 kDa glycoprotein common to platelets and endothelial cells: heterogeneity in antibody binding to human aortic endothelial cells. *Thromb Haemostasis* 1991 Oct 1;66(4):494-9.
2. Yoshino N, Ami Y, Terao K, Tashiro F, Honda M. Upgrading of flow cytometric analysis for absolute counts, cytokines and other antigenic molecules of cynomolgus monkeys (*Macaca fascicularis*) by using anti-human cross-reactive antibodies. *Exp Anim* 2000 Apr;49(2):97-110.
3. Ciarlet M, Crawford SE, Cheng E, Blutt SE, Rice DA, Bergelson JM, et al. VLA-2 (alpha2beta1) integrin promotes rotavirus entry into cells but is not necessary for rotavirus attachment. *J Virol* 2002 Feb;76(3):1109-23.

## 7. EXPLANATION OF SYMBOLS

	Form
	Catalog reference
	Contains sufficient for <n> test
	Quantity per test
	Regulatory Status
	Research Use Only
	Manufacturer

## 8. MANUFACTURED BY:



### IMMUNOSTEP S.L.

**Address:** Avda. Universidad de Coimbra, s/n  
 Cancer Research Center (C.I.C)  
 Campus de Unamuno  
 37007 Salamanca (Spain)

**Tel./fax:** (+34) 923 294 827  
**E-mail:** [info@immunostep.com](mailto:info@immunostep.com)  
[www.immunostep.com](http://www.immunostep.com)